

Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Bioanalytical Standard
Product Model Number 1530
Lot Number 18K100150
Method Comparative enzymatic testing of Production standard vs Metrology standard

| Chemistry Tested | Determination | Tolerance | Result |
|------------------|---|----------------------|----------|
| LACTATE | (Average of Production)-(Average of Metrology) | $\pm 0.05\text{g/L}$ | 0.00 g/L |
| | Coefficient of variation for production standard (CV,%) | $\leq 2.0\%$ | 0.3 % |

Conforms to acceptance criteria

Completed by: Kristin Stover

Date: 10/10/2018



A handwritten signature in blue ink, appearing to read 'Lisa M. Abel'.

Lisa M. Abel
Director, Quality and Corporate
Responsibility