

# Certificate of Analysis



**YSI Incorporated**

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category            Membrane  
Product Model Number    2329  
Lot Number                 19B100483  
Method                      Amperometric testing with standards

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	6.0-30.0 nA	19.7 nA
	Average Linearity	14.3-15.7mmol/L	14.89 mmol/L
	Potassium Ferrocyanide	-0.3 - 0.3mmol/L	0.03 mmol/L

Conforms to acceptance criteria

Completed by:            Christy Woodford/Sandy Crawford            2/20/2019  
Your Name

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



A handwritten signature in blue ink that reads 'Lisa M. Abel'.

Lisa M. Abel  
Director, Quality and Corporate  
Responsibility