

# Certificate of Analysis



**YSI Incorporated**

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Bioanalytical Standard  
Product Model Number 2356  
Lot Number 19J100705  
Method Comparative enzymatic testing of Production standard vs Metrology standard

| Chemistry Tested | Determination   | Tolerance | Result  |
|------------------|---|-----------|---------|
| GLUCOSE          | (Average of Production)-(Average of Metrology)          | ±25mg/L   | 11 mg/L |
|                  | Coefficient of variation for production standard (CV,%) | ≤2.0%     | 0.7 %   |

Conforms to acceptance criteria

Completed by: Lucas Stover

Date: 10/1/2019



A handwritten signature in blue ink, appearing to read 'Lisa M. Abel'.

Lisa M. Abel  
Director, Quality and Corporate  
Responsibility