

Certificate of Analysis



YSI Incorporated
Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA
Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Bioanalytical Standard
Product Model Number 2792
Lot Number 20B100032
Method Comparative enzymatic testing of Production standard vs. Metrology standard

Chemistry Tested	Determination	Tolerance	Result
ETHANOL STD A 1.0G/L	(Average of production) - (average of metrology)	±20mg/L	2.5mg/L
	Coefficient of variation for production standard (CV, %)	≤2%	0.5%
ETHANOL STD B 0.5G/L	(Average of production) - (average of metrology)	±10mg/L	2 mg/L
	Coefficient of variation for production standard (CV, %)	≤2%	0.5%

Conforms to acceptance criteria

Completed by: Lucas Stover Date: 02/06/20



Lisa M. Abel
Director, Quality and Corporate
Responsibility