

Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Membrane
Product Model Number 2329
Lot Number 20H100418
Method Amperometric testing with standards

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	6.0-30.0 nA	13.2 nA
	Average Linearity	14.3-15.7mmol/L	14.88 mmol/L
	Potassium Ferrocyanide	-0.3 - 0.3mmol/L	0.01 mmol/L

Conforms to acceptance criteria

Completed by: _____ Christy Woodford _____ 8/26/2020
Your Name

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



A handwritten signature in blue ink that reads "Gregory W. Popp".

Gregory W. Popp
Quality Manager Analytics