

# Certificate of Analysis



**YSI Incorporated**

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Bioanalytical Standard  
Product Model Number 2747  
Lot Number 20L100416  
Method Comparative enzymatic testing of Production standard vs Metrology standard

Chemistry Tested	Determination	Tolerance	Result
GLUCOSE	(Average of Production)-(Average of Metrology)	$\pm 0.01$ g/L	0.01 g/L
	Coefficient of variation for production standard (CV,%)	$\leq 2.0\%$	1.1 %
LACTATE	(Average of Production)-(Average of Metrology)	$\pm 0.01$ g/L	0.00 g/L
	Coefficient of variation for production standard (CV,%)	$\leq 2.0\%$	0.9 %

Conforms to acceptance criteria

Completed by: Amanda Fenner

Date: 11/19/2020



A handwritten signature in blue ink, appearing to read 'Lisa M. Abel'.

Lisa M. Abel  
Director, Quality and Corporate  
Responsibility