

Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Membrane
Product Model Number 2725
Lot Number 21H100511
Method Amperometric testing with standards

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	5.0-25.0 nA	5.76
	Average Linearity	2.375-2.625 g/L	2.56
	Potassium Ferrocyanide	-0.2-0.2 g/L	0.00
Final	Average Calibration Sensitivity	5.0-25.0 nA	5.79
	Average Linearity	2.375-2.625 g/L	2.47
	Potassium Ferrocyanide	-0.2-0.2 g/L	0.00

Conforms to acceptance criteria

Completed by: _____ Christy Woodford _____ 8/30/2021
Your Name



A handwritten signature in blue ink that reads 'Gregory W. Popp'.

Gregory W. Popp
Quality Manager