

# Certificate of Analysis



**YSI Incorporated**

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

---

Product Category	Bioanalytical Standard
Product Model Number	2747
Lot Number	22A100687
Method	Comparative enzymatic testing of Production standard vs Metrology standard

---

Chemistry Tested	Determination	Tolerance	Result
GLUCOSE	(Average of Production)-(Average of Metrology)	$\pm 0.01\text{g/L}$	0.00 g/L
	Coefficient of variation for production standard (CV,%)	$\leq 2.0\%$	0.77 %
LACTATE	(Average of Production)-(Average of Metrology)	$\pm 0.01\text{g/L}$	0.00 g/L
	Coefficient of variation for production standard (CV,%)	$\leq 2.0\%$	0.69 %

---

Conforms to acceptance criteria

Completed by: Nick Rittenhouse

Date: 2/3/2022

---

*The procedures used to manufacture and test this product have been verified by QA for GMP compliance.*



A handwritten signature in blue ink that reads 'Gregory W. Popp'.

Gregory W. Popp  
Quality Manager, Analytics