

# Certificate of Analysis



**YSI Incorporated**

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Bioanalytical Standard  
Product Model Number 2801  
Lot Number 22F100405  
Method Comparative enzymatic testing of Production standard vs Metrology standard

| Chemistry Tested | Determination   | Tolerance      | Result    |
|------------------|---|----------------|-----------|
| GLUCOSE          | (Average of Production)-(Average of Metrology)          | $\pm 5.0$ mg/L | 0.15 mg/L |
|                  | Coefficient of variation for production standard (CV,%) | $\leq 2.0\%$   | 0.6 %     |

Conforms to acceptance criteria

Completed by: Nick Rittenhouse

Date: 6/21/2022

*The procedures used to manufacture and test this product have been verified by QA for GMP compliance.*



A handwritten signature in blue ink that reads 'Gregory W. Popp'.

Gregory W. Popp  
Quality Manager, Analytics