



Certificate of Analysis
Bioanalytical Standard
2356

1725 Brannum Lane
Yellow Springs, Ohio
45387-1107 USA
+1 937-767-7241

Method:

Comparative enzymatic testing of Production standard vs Metrology standard

Testing Results

Lot Number: 22K100392

Chemistry Tested	Determination	Tolerance	Result
GLUCOSE	(Average of Production)- (Average of Metrology)	±25 mg/L	10.85 mg/L
	Coefficient of variation for production standard (CV,%)	≤2.0%	0.8 %

Conforms to acceptance criteria

Completed by: afenner Date: 10/20/2022

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.