

Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Membrane
Product Model Number 2735
Lot Number 23C100091
Method Amperometric testing with standards

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	8.0-40.0 nA	13.2 nA
	Average Linearity	7.6-8.4mmol/L	7.93 mmol/L
	Potassium Ferrocyanide	< 0.4mmol/L	0.03 mmol/L

Conforms to acceptance criteria

Completed by:

Christy Woodford

3/7/2023

Your Name



A handwritten signature in blue ink that reads 'Gregory W. Popp'.

Gregory W. Popp
Quality Manager Analytics