Certificate of Analysis

YSI Incorporated
Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA
Phone 937 767-7241•800 765-4974•Fax 937 767-9353

Product Category	Membrane
Product Model Number	2725
Lot Number	23C100093
Method	Amperometric testing with standards

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	5.0-25.0 nA	6.27 nA
	Average Linearity	2.375-2.625 g/L	2.470 g/L
	Potassium Ferrocyanide	-0.2-0.2 g/L	0.00 g/L
Final	Average Calibration Sensitivity	5.0-25.0 nA	6.75 nA
	Average Linearity	2.375-2.625 g/L	2.490 g/L
	Potassium Ferrocyanide	-0.2-0.2 g/L	0.00 g/L

Conforms to acceptance criteria

Completed by: Christy Woodford 3/13/2023
Your Name

ISO 9001 CERTIFIED
YSI INCORPORATED

Gregory W. Popp
Quality Manager