



# Certificate of Analysis

1725 Brannum Lane  
Yellow Springs, Ohio  
45387-1107 USA  
+1 937-767-7241

Product Category            ISE Reagent  
Product Model Number    2971 (1000 mg/L Potassium, 10 mg/L Ammonium)  
Lot Number                 23L100547  
Method                      Comparative ISE testing of Production reagent vs. Metrology reagent

## Testing Results

Chemistry Tested	Determination	Tolerance	Result
Potassium	(Average of Production) – (Average of Metrology)	$\leq 20$ mg/L (2%)	0.34
	Coefficient of Variation (CV) for Production reagent	$\leq 3\%$	1.64

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.

Technician

Date

Nick Rittenhouse

12/1/2023