

Certificate of Analysis



YSI Incorporated
Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA
Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

a xylem brand

Product Category Membrane
Product Model Number 2761
Lot Number 24A100210
Method Amperometric testing with standards

Time of Testing	Determination	Specification	Result
Initial	Average calibration sensitivity	10.0 – 60.0 nA	21.67 nA
	Average Linearity	27.0 – 33.0 g/L	39.38 g/L
	Potassium Ferrocyanide interference	0 – 0.07 g/L	0.0 g/L
Day = _____	Average calibration sensitivity	N/A	N/A
	Average Linearity	N/A	N/A
	Potassium Ferrocyanide interference	N/A	N/A

Conforms to acceptance criteria

Completed by:

Christy Woodford
Christy Woodford

Date:

1/31/2024

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



Gregory W. Popp

Gregory W. Popp
Quality Manager, Analytics