

Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

| | |
|----------------------|-------------------------------------|
| Product Category | Membrane |
| Product Model Number | 2329 |
| Lot Number | 16L100817 |
| Method | Amperometric testing with standards |

| Time of Testing | Determination | Specification | Results |
|-----------------|---------------------------------|------------------|--------------|
| Initial | Average Calibration Sensitivity | 10.0-30.0 nA | 16.0 nA |
| | Average Linearity | 14.3-15.7mmol/L | 14.85 mmol/L |
| | Potassium Ferrocyanide | -0.3 - 0.3mmol/L | 0.02 mmol/L |

Conforms to acceptance criteria

Completed by: Christy Woodford 12/8/2016



A handwritten signature in blue ink, appearing to read 'Lisa M. Abel'.

Lisa M. Abel
Director, Quality and Corporate
Responsibility