Certificate of Analysis

YSI Incorporated
Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA
Phone 937 767-7241•800 765-4974•Fax 937 767-9353

Product Category	Membrane
Product Model Number	2735
Lot Number	18D100190
Method	Amperometric testing with standards

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	8.0-40.0 nA	15.8 nA
	Average Linearity	7.6-8.4mmol/L	8.00 mmol/L
	Potassium Ferrocyanide	< 0.4mmol/L	0.00 mmol/L

Conforms to acceptance criteria

Completed by: Amanda DuVall 4/10/2018
Your Name

Lisa M. Abel
Director of Quality