

Certificate of Analysis



YSI Incorporated
Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA
Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

a xylem brand

| | |
|----------------------|---|
| Product Category | Buffer and solid products |
| Product Model Number | <u>1579</u> |
| Lot Number | <u>20K100459</u> |
| Method | Verification of mass and pH (for buffers) |

| Determination | Specification | Result |
|---------------------------|---------------|--------|
| Average mass | N/A | |
| Average pH (buffers only) | 9.00 ± 0.10 | 8.96 |

Conforms to acceptance criteria

Completed by: Nick Rittenhouse Date: 10/20/2020
Amanda Fenner

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



Gregory W. Popp

Gregory W. Popp
Quality Manager, Analytics