

# Certificate of Analysis



**YSI Incorporated**

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Bioanalytical Standard  
Product Model Number 2726  
Lot Number 21E100821  
Method Comparative enzymatic testing of Production standard vs Metrology standard

| Chemistry Tested | Determination   | Tolerance | Result  |
|------------------|---|-----------|---------|
| METHANOL PART A  | (Average of Production)-(Average of Metrology)          | ±20mg/L   | 18 mg/L |
|                  | Coefficient of variation for production standard (CV,%) | ≤2.0%     | 0.6 %   |
| METHANOL PART B  | (Average of Production)-(Average of Metrology)          | ±50mg/L   | 11 mg/L |
|                  | Coefficient of variation for production standard (CV,%) | ≤2.0%     | 1.2 %   |

Conforms to acceptance criteria

Completed by: Nick Rittenhouse Date: 6/3/2021

*The procedures used to manufacture and test this product have been verified by QA for GMP compliance.*



A handwritten signature in blue ink that reads 'Gregory W. Popp'.

Gregory W. Popp  
Quality Manager, Analytics