

# Certificate of Analysis



**YSI Incorporated**

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category            Membrane  
Product Model Number    2329  
Lot Number                 22C100318  
Method                      Amperometric testing with standards

| Time of Testing | Determination                   | Specification    | Results      |
|-----------------|---------------------------------|------------------|--------------|
| Initial         | Average Calibration Sensitivity | 6.0-30.0 nA      | 13.1 nA      |
|                 | Average Linearity               | 14.3-15.7mmol/L  | 14.84 mmol/L |
|                 | Potassium Ferrocyanide          | -0.3 - 0.3mmol/L | 0.01 mmol/L  |

Conforms to acceptance criteria

Completed by:            Sandy Crawford  
                                  Your Name

3/14/2022

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



A handwritten signature in blue ink that reads 'Gregory W. Popp'.

Gregory W. Popp  
Quality Manager Analytics