

Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category	Membrane
Product Model Number	2365
Lot Number	22K100397
Method	Amperometric testing with standards

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	8.0-45.0 nA	09.0 nA
	Average Linearity	8.55-9.45 g/L	g/L
	Potassium Ferrocyanide	0.00-0.05	#DIV/0!

Conforms to acceptance criteria

Completed by: Sandy Crawford
Your Name

10/14/2022



A handwritten signature in blue ink that reads 'Gregory W. Popp'.

Gregory W. Popp
Quality Manager Analytics