



Certificate of Analysis

Bioanalytical Standard

2747

1725 Brannum Lane
Yellow Springs, Ohio
45387-1107 USA
+1 937-767-7241

Method:

Comparative enzymatic testing of Production standard vs Metrology standard

Testing Results

Lot Number: 22L100049

Chemistry Tested	Determination	Tolerance	Result
GLUCOSE	(Average of Production)- (Average of Metrology)	±0.01 g/L	0.001 g/L
	Coefficient of variation for production standard (CV,%)	≤2.0%	1.01 %
LACTATE	(Average of Production)- (Average of Metrology)	±0.01 g/L	0.001 g/L
	Coefficient of variation for production standard (CV,%)	≤2.0%	0.98 %

Conforms to acceptance criteria

Completed by: NRITTEN Date: 11/8/2022

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.