

# Certificate of Analysis



**YSI Incorporated**  
Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA  
Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

a xylem brand

Product Category            Buffer and solid products  
Product Model Number      2357  
Lot Number                    230104GFS098005  
Method                         Verification of mass and pH (for buffers)

Determination	Specification	Result
Average mass	6.00 – 6.70 g	6.17 g
Average pH (buffers only)	7.1-7.5	7.3

Conforms to acceptance criteria

Completed by: Carly Garneau

A handwritten signature in cursive script, appearing to read 'Carly Garneau'.

Date: 01/23/2023

A handwritten date in cursive script, appearing to read '23-JAN-2023'.

*The procedures used to manufacture and test this product have been verified by QA for GMP compliance.*



A handwritten signature in cursive script, appearing to read 'Gregory W. Popp'.

Gregory W. Popp  
Quality Manager, Analytics