

Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Membrane
Product Model Number 2786
Lot Number 23C100585
Method Amperometric testing with standards

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	5 - 25 nA	7.68
	Average Linearity	3.04 - 3.36 g/L	3.22
	Potassium Ferrocyanide	-0.2 - 0.2 g/L	0.00
Final	Average Calibration Sensitivity	5 - 25 nA	7.14
	Average Linearity	3.04 - 3.36 g/L	3.18
	Potassium Ferrocyanide	-0.2 - 0.2 g/L	0.00

Conforms to acceptance criteria

Completed by: Christy Woodford

4/11/2023

Your Name



A handwritten signature in blue ink that reads 'Gregory W. Popp'.

Gregory W. Popp
Quality Manager Analytics