



# Certificate of Analysis

1725 Brannum Lane  
Yellow Springs, Ohio  
45387-1107 USA  
+1 937-767-7241

Product Category            ISE Reagent  
Product Model Number    7179 (100 mg/L Ammonium, 200 mg/L Potassium)  
Lot Number                 23G100041  
Method                      Comparative ISE testing of Production reagent vs. Metrology reagent

## Testing Results

Chemistry Tested	Determination	Tolerance	Result
Ammonium	(Average of Production) – (Average of Metrology)	≤ 2 mg/L (2%)	0.62
	Coefficient of Variation (CV) for Production reagent	≤ 3%	0.15
Potassium	(Average of Production) – (Average of Metrology)	≤ 4 mg/L (2%)	2.37
	Coefficient of Variation (CV) for Production reagent	≤ 3%	0.87

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.

Technician

Date

Amanda Fenner

7/10/2023