

Certificate of Analysis



YSI Incorporated
Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA
Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

a **xylem** brand

Product Category Buffer and solid products
Product Model Number 1579
Lot Number 23K100346
Method Verification of mass and pH (for buffers)

Determination	Specification	Result
Average mass	N/A	N/A
Average pH (buffers only)	9.00 ± 0.10	9.00

Conforms to acceptance criteria

Completed by: Nick Rittenhouse Date: 10/19/2023

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



Gregory W. Popp
Quality Manager, Analytics