

Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Sucrose Membrane
Product Model Number 2703
Lot Number 23K100448
Method Amperometric testing with standards

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	10.0-40.0 nA	10.7 nA
	Average Linearity	23.75-26.25 g/L	25.40 g/L
	Potassium Ferrocyanide Interference	0.0-0.3	0.00 g/L

Conforms to acceptance criteria

Completed by: Christy Woodford
Your Name

Date: 10/24/2023

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



A handwritten signature in blue ink that reads 'Gregory W. Popp'.

Gregory W. Popp
Quality Manager Analytics