

# Certificate of Analysis



**YSI Incorporated**  
Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA  
Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

a xylem brand

Product Category                      Membrane  
Product Model Number                      2702  
Lot Number                                      24A100580  
Method    Amperometric testing with standards

Time of Testing	Determination	Specification	Result
Initial	Average calibration sensitivity	10 - 30 nA	14.82 nA
	Average Linearity	23.75 – 26.25 g/L	25.83 g/L
	Potassium Ferrocyanide interference	N/A	N/A
Day = _____	Average calibration sensitivity	N/A	N/A
	Average Linearity	N/A	N/A
	Potassium Ferrocyanide interference	N/A	N/A

Conforms to acceptance criteria

Completed by:

  
Christy Woodford

Date:

1/31/2024

*The procedures used to manufacture and test this product have been verified by QA for GMP compliance.*





Gregory W. Popp  
Quality Manager, Analytics