

Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Membrane
Product Model Number 2365
Lot Number 17G100650
Method Amperometric testing with standards

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	8.0-45.0 nA	12.4 nA
	Average Linearity	8.55-9.45 g/L	9.02 g/L
	Potassium Ferrocyanide	0.00-0.05	0.00 g/L

Conforms to acceptance criteria

Completed by: Christy Woodford
 Your Name

8/9/2017



A handwritten signature in blue ink, appearing to read 'Lisa M. Abel'.

Lisa M. Abel
Director, Quality and Corporate
Responsibility