

Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Membrane
Product Model Number 2365
Lot Number 17J100028
Method Amperometric testing with standards

| Time of Testing | Determination | Specification | Results |
|-----------------|---------------------------------|---------------|----------|
| Initial | Average Calibration Sensitivity | 8.0-45.0 nA | 09.7 nA |
| | Average Linearity | 8.55-9.45 g/L | 9.09 g/L |
| | Potassium Ferrocyanide | 0.00-0.05 | 0.00 g/L |

Conforms to acceptance criteria

Completed by: Amanda Mills 09/13/2017

Your Name



A handwritten signature in blue ink, appearing to read 'Lisa M. Abel'.

Lisa M. Abel
Director, Quality and Corporate
Responsibility