Certificate of Analysis

YSI Incorporated
Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA
Phone 937 767-7241•800 765-4974•Fax 937 767-9353

Product Category	Membrane
Product Model Number	2329
Lot Number	18C101075
Method	Amperometric testing with standards

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	6.0-30.0 nA	15.4 nA
	Average Linearity	14.3-15.7mmol/L	15.04 mmol/L
	Potassium Ferrocyanide	-0.3 - 0.3mmol/L	0.00 mmol/L

Conforms to acceptance criteria

Completed by: Amanda DuVall 04/5/2018
Your Name

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.

ISO 9001 CERTIFIED YSI INCORPORATED

Lisa M. Abel
Director, Quality and Corporate Responsibility