

# Certificate of Analysis



**YSI Incorporated**

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category            Membrane  
Product Model Number    2365  
Lot Number                 18K100142  
Method                      Amperometric testing with standards

| Time of Testing | Determination                   | Specification | Results  |
|-----------------|---------------------------------|---------------|----------|
| Initial         | Average Calibration Sensitivity | 8.0-45.0 nA   | 11.6 nA  |
|                 | Average Linearity               | 8.55-9.45 g/L | 9.15 g/L |
|                 | Potassium Ferrocyanide          | 0.00-0.05     | 0.00 g/L |

Conforms to acceptance criteria

Completed by:

Sandy Crawford

10/10/2018

Your Name



A handwritten signature in blue ink, appearing to read 'Lisa M. Abel'.

Lisa M. Abel  
Director, Quality and Corporate  
Responsibility