

Certificate of Analysis



YSI Incorporated
Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA
Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Buffer and solid products
Product Model Number 2357
Lot Number 200103GFS098005
Method Verification of mass and pH (for buffers)

Determination	Specification	Result
Average mass	6.00 – 6.70 g	6.27g
Average pH (buffers only)	7.1-7.5	7.3

Conforms to acceptance criteria

Completed by: Lucas Stover Date: 02/05/2020



Lisa M. Abel
Director, Quality and Corporate
Responsibility