

# Certificate of Analysis



**YSI Incorporated**  
Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA  
Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

a xylem brand

Product Category                      Bioanalytical Standard  
Product Model Number                      2790  
Lot Number                                      21B100431  
Method    Comparative enzymatic testing of Production standard vs. Metrology standard

Chemistry Tested	Determination	Tolerance	Result
Ethanol, 2.0 g/L	(Average of production) - (average of metrology)	$\pm 40$ mg/L	20.85 mg/L
	Coefficient of variation for production standard (CV, %)	$\leq 2.0\%$	0.71%
Ethanol, 3.2 g/L	(Average of production) - (average of metrology)	$\pm 64$ mg/L	1.15 mg/L
	Coefficient of variation for production standard (CV, %)	$\leq 2.0\%$	0.75%

Conforms to acceptance criteria

Completed by: Nick Rittenhouse                                      Date: 2/19/2021

*The procedures used to manufacture and test this product have been verified by QA for GMP compliance.*



A handwritten signature in blue ink that reads 'Gregory W. Popp'.

Gregory W. Popp  
Quality Manager, Analytics