

Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

| | |
|----------------------|--|
| Product Category | Bioanalytical Standard |
| Product Model Number | 2755 |
| Lot Number | 21C100746 |
| Method | Comparative enzymatic testing of Production standard vs Metrology standard |

| Chemistry Tested | Determination | Tolerance | Result |
|------------------|---|-------------|------------|
| GLUTAMATE | (Average of Production)-(Average of Metrology) | ±0.05mmol/L | 0.04mmol/L |
| | Coefficient of variation for production standard (CV,%) | ≤2.0% | 0.9 % |

Conforms to acceptance criteria

Completed by: Nick Rittenhouse

Date: 3/31/2021

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



A handwritten signature in blue ink that reads 'Gregory W. Popp'.

Gregory W. Popp
Quality Manager, Analytics