

Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Membrane
 Product Model Number 2725
 Lot Number 21F100557
 Method Amperometric testing with standards

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	5.0-25.0 nA	6.17
	Average Linearity	2.375-2.625 g/L	
	Potassium Ferrocyanide	-0.2-0.2 g/L	0.00
Final	Average Calibration Sensitivity	5.0-25.0 nA	7.35
	Average Linearity	2.375-2.625 g/L	
	Potassium Ferrocyanide	-0.2-0.2 g/L	0.00

Conforms to acceptance criteria

Completed by: _____ Christy Woodford 6/28/2021
 Your Name



Gregory W. Popp

Gregory W. Popp
 Quality Manager