

Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Bioanalytical Standard
Product Model Number 2792
Lot Number 21J100432
Method Comparative enzymatic testing of Production standard vs Metrology standard

Chemistry Tested	Determination	Tolerance	Result
ETHANOL PART A	(Average of Production)-(Average of Metrology)	±20mg/L	2.35 mg/L
	Coefficient of variation for production standard (CV,%)	≤2.0%	0.48 %
ETHANOL PART B	(Average of Production)-(Average of Metrology)	±10mg/L	3.65 mg/L
	Coefficient of variation for production standard (CV,%)	≤2.0%	0.57 %

Conforms to acceptance criteria

Completed by: Nick Rittenhouse

Date: 9/22/2021

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



A handwritten signature in blue ink that reads 'Gregory W. Popp'.

Gregory W. Popp
Quality Manager, Analytics