

Certificate of Analysis



YSI Incorporated
Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA
Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

a xylem brand

Product Category	Buffer and solid products
Product Model Number	<u>1579</u>
Lot Number	<u>21K100032</u>
Method	Verification of mass and pH (for buffers)

Determination	Specification	Result
Average mass	N/A	
Average pH (buffers only)	9.00 ± 0.10	8.96

Conforms to acceptance criteria

Completed by: Amanda Fenner Date: 10/05/2021

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



Gregory W. Popp
Quality Manager, Analytics