

Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Bioanalytical Standard
Product Model Number 2785
Lot Number 22F100021
Method Comparative enzymatic testing of Production standard vs Metrology standard

| Chemistry Tested | Determination | Tolerance | Result |
|------------------|---|-----------|----------|
| GLUCOSE | (Average of Production)-(Average of Metrology) | ±0.13g/L | 0.01 g/L |
| | Coefficient of variation for production standard (CV,%) | ≤2.0% | 1.27 % |
| GALACTOSE | (Average of Production)-(Average of Metrology) | ±0.13g/L | 0.05 g/L |
| | Coefficient of variation for production standard (CV,%) | ≤2.0% | 1.12 % |

Conforms to acceptance criteria

Completed by: Amanda Fenner

Date: 6/2/2022

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



A handwritten signature in blue ink that reads 'Gregory W. Popp'.

Gregory W. Popp
Quality Manager, Analytics