

# Certificate of Analysis



**YSI Incorporated**

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category      Sucrose Membrane  
Product Model Number    2703  
Lot Number              22H100305  
Method                  Amperometric testing with standards

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	10.0-40.0 nA	10.7 nA
	Average Linearity	23.75-26.25 g/L	24.89 g/L
	Potassium Ferrocyanide Interference	0.0-0.3	0.01 g/L

Conforms to acceptance criteria

Completed by: Christy Woodford  
Your Name

Date: 8/24/2022

*The procedures used to manufacture and test this product have been verified by QA for GMP compliance.*



A handwritten signature in blue ink that reads 'Gregory W. Popp'.

Gregory W. Popp  
Quality Manager Analytics