

Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

| | |
|----------------------|-------------------------------------|
| Product Category | Membrane |
| Product Model Number | 2771 |
| Lot Number | 22L100113 |
| Method | Amperometric testing with standards |

| Time of Testing | Determination | Specification | Results |
|-----------------|-------------------------------------|----------------|------------|
| Initial | Average Calibration Sensitivity | >5 nA | 11.83nA |
| | Average Linearity | 427 - 473 mg/L | 451.67mg/L |
| | Potassium Ferrocyanide Interference | <15 mg/L | 0.00mg/L |
| Day 2 | Average Calibration Sensitivity | >5 nA | 13.56nA |
| | Average Linearity | 427 - 473 mg/L | 433.67mg/L |
| | Potassium Ferrocyanide Interference | <15 mg/L | 0.13mg/L |

Conforms to acceptance criteria

Completed by:

Christy Woodford

11/9/2022

Your Name



A handwritten signature in blue ink that reads 'Gregory W. Popp'.

Gregory W. Popp
Quality Manager Analytics