

Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Bioanalytical Standard
Product Model Number 2726
Lot Number 23L100144
Method Comparative enzymatic testing of Production standard vs Metrology standard

Chemistry Tested	Determination	Tolerance	Result
METHANOL PART A	(Average of Production)-(Average of Metrology)	±20mg/L	10 mg/L
	Coefficient of variation for production standard (CV,%)	≤2.0%	0.7 %
METHANOL PART B	(Average of Production)-(Average of Metrology)	±50mg/L	32 mg/L
	Coefficient of variation for production standard (CV,%)	≤2.0%	1.5 %

Conforms to acceptance criteria

Completed by: Amanda Fenner

Date: 11/9/2023

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



A handwritten signature in blue ink that reads 'Gregory W. Popp'.

Gregory W. Popp
Quality Manager, Analytics