Certificate of Analysis

YSI Incorporated
Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA
Phone 937 767-7241•800 765-4974•Fax 937 767-9353

a xylem brand

Product Category: Buffer and solid products

Product Model Number: 2705

Lot Number: 240501GFS002699

Method: Verification of mass and pH (for buffers)

Determination	Specification	Result
Average mass	6.35 ± 0.35g	6.32g
Average pH (buffers only)	7.3 ± 0.2	7.3

The procedures used to manufacture and test this product have been verified by QA for compliance. The electronic signature below is representative of the Chemistry Technician testing and verifying conformity to acceptance criteria.

Completed by: Amanda Fenner Date: 5/06/2024

ISO 9001 CERTIFIED YSI INCORPORATED