

Certificate of Analysis



YSI Incorporated
Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA
Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

a xylem brand

Product Category Membrane
Product Model Number 2702
Lot Number 25C100409
Method Amperometric testing with standards

Time of Testing	Determination	Specification	Result
Initial	Average calibration sensitivity	10 - 30 nA	11.09 nA
	Average Linearity	23.75 - 26.25 g/L	25.03 g/L
	Potassium Ferrocyamide interference	N/A	N/A
Day =	Average calibration sensitivity	N/A	N/A
	Average Linearity	N/A	N/A
	Potassium Ferrocyamide interference	N/A	N/A

The procedures used to manufacture and test this product have been verified by QA for compliance. The electronic signature below is representative of the Chemistry Technician testing and verifying conformity to acceptance criteria.

Completed by: Christy Woodford Date: 3/26/2025

A handwritten signature in cursive script, appearing to read 'Christy Woodford', is written over a horizontal line.

