

# Certificate of Analysis



**YSI Incorporated**

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

a xylem brand

Product Category                      Buffer and solid products  
Product Model Number                1579  
Lot Number                                25D100509  
Method                                      Verification of mass and pH (for buffers)

| Determination             | Specification | Result |
|---------------------------|---------------|--------|
| Average mass              | N/A           | N/A    |
| Average Ph (buffers only) | 9.00 ± 0.10   | 8.97   |

*The procedures used to manufacture and test this product have been verified by QA for compliance. The electronic signature below is representative of the Chemistry Technician testing and verifying conformity to acceptance criteria.*

Completed by:      Amanda Fenner                       Date:      4/25/2025      4/25/2025

