



# Certificate of Analysis

1725 Brannum Lane  
Yellow Springs, Ohio  
45387-1107 USA  
+1 937-767-7241

Product Category ISE Reagent  
Product Model Number 2970 (10 mg/L Ammonium, 10 mg/L Potassium)  
Lot Number 26A100115  
Method Comparative ISE testing of Production reagent vs. Metrology reagent

## Testing Results

Chemistry Tested	Determination	Tolerance	Result
Ammonium	(Average of Production) – (Average of Metrology)	≤ 0.2 mg/L (2%)	0.05
	Coefficient of Variation (CV) for Production reagent	≤ 5%	0.20
Potassium	(Average of Production) – (Average of Metrology)	≤ 0.3 mg/L (3%)	0.13
	Coefficient of Variation (CV) for Production reagent	≤ 5%	1.73

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.

Technician

Amanda Fenner

Date

1/13/2026