

# Certificate of Analysis



## YSI Incorporated

**Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA**

Phone 937 767-7241•800 765-4974•Fax 937 767-9353

Product Category	Membrane
Product Model Number	7140
Lot Number	26A100481
Method	Amperometric testing with standards

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	8 - 75 nA	14.76 nA
	Average Linearity	38 - 42 g/L	40.20 g/L

*The procedures used to manufacture and test this product have been verified by QA for compliance. The electronic signature below is representative of the Chemistry Technician testing and verifying conformity to acceptance criteria.*

Completed by: Christy Woodford  
Your Name

Date: 1/21/2026

