



# ***Certificate of Analysis***

1725 Brannum Lane  
Yellow Springs, Ohio  
45387-1107 USA  
+1 937-767-7241

Product Category ISE Reagent  
Product Model Number 2972 (500 mg/L Ammonium, 10 mg/L Potassium)  
Lot Number 26A100514  
Method Comparative ISE testing of Production reagent vs. Metrology reagent

## **Testing Results**

<b>Chemistry Tested</b>	<b>Determination</b>	<b>Tolerance</b>	<b>Result</b>
Ammonium	(Average of Production) – (Average of Metrology)	≤ 10 mg/L (2%)	0.61
	Coefficient of Variation (CV) for Production reagent	≤ 3%	0.55

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.

Technician

Amanda Fenner

Date

1/20/2026