

# Certificate of Analysis



**YSI Incorporated**  
Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA  
Phone 937 767-7241•800 765-4974•Fax 937 767-9353

a **xylem** brand

Product Category	Check Solution
Product Model Number	<u>7173</u>
Lot Number	<u>26A100516</u>
Method	Comparative conductivity testing of Production standard vs. Metrology standard

Chemistry Tested	Determination	Tolerance	Result
Sodium	(Average of production) – (average of metrology)	± 1.0 %	0.40 %
	Coefficient of variation for production standard (CV, %)	≤ 2.0 %	0.82 %

The procedures used to manufacture and test this product have been verified by QA for compliance. The electronic signature below is representative of the Chemistry Technician testing and verifying conformity to acceptance criteria.

Completed by: Amanda Fenner Date: 1/23/2026

