



# North Central Laboratories

P.O. Box 8, Birnamwood, WI 54414

## Certificate of Analysis

Product Number: YSI-3821  
Product Name: Buffer Solution pH 4.00  
Lot Number: 25L1R  
Expiration Date: 11/12/27

The above material is certified to have a pH of  $4.00 \pm 0.02$  at  $25^{\circ}\text{C}$ , traceable to NIST reference materials.

*Mark Mahoney*

Mark Mahoney  
Chemist  
NCL of Wisconsin Inc.

11/14/25

Date

The above material has been tested and meets the stated specifications after manufacture. Conditions of storage and use are beyond our knowledge and control. The possibility of contamination by the end user exists. Stated expiration date is based on our experience with retained samples, and is not to be construed as a warranty.

1.800.648.7836

Fax: (715) 449.2454

[www.nclabs.com](http://www.nclabs.com)

Email: [nclabs@nclabs.com](mailto:nclabs@nclabs.com)