



North Central Laboratories

P.O. Box 8, Birnamwood, WI 54414

Certificate of Analysis

Product Number: YSI-3822
Product Name: Buffer Solution pH 7.00
Lot Number: 25K2S
Expiration Date: 10/29/27

The above material is certified to have a pH of 7.00 \pm 0.02 at 25°C, traceable to NIST reference materials.

Mark Mahoney

Mark Mahoney
Chemist
NCL of Wisconsin Inc.

10/31/25

Date

The above material has been tested and meets the stated specifications after manufacture. Conditions of storage and use are beyond our knowledge and control. The possibility of contamination by the end user exists. Stated expiration date is based on our experience with retained samples, and is not to be construed as a warranty.

1.800.648.7836

Fax: (715) 449.2454

www.nclabs.com

Email: nclabs@nclabs.com